

A-FLX™ FFPE Cell Pellet Array Stimulated PBMC Array (PMA/Ionomycin)

GENERAL INFORMATION

Product Name:	A-FLX™ FFPE Cell Pellet Array Stimulated PBMC Array (PMA/Ionomycin)
Reference Number:	3600-0310 Block 3600-0320 Slide (4µm)
Date of Manufacturing:	See product label
Lot Number:	See product label
Intended Use:	For research use only

SAFETY AND PRECAUTIONS

This product does not contain hazardous material. Wear appropriate personal protective equipment (PPE) when handling reagents and biological specimens.

DESCRIPTION


Cell Line:	PBMC
Tissue of Origin:	Blood
Fixation Condition:	10% neutral buffered formalin (NBF) for 24 hours at 24-27°C
Product Format:	One unstained section mounted on Superfrost™ Plus slide. Section thickness: 4µm

RECOMMENDED PROCEDURES

Staining using FFPE slides:

1. Bake slides at 60°C for 30-60 min.
2. Deparaffinize 3 times in Xylene or Xylene substitute for 5 min each time
3. Rinse 2 times in 100% ethanol for 1 min each.
4. Rehydrate in ethanol series (95% 1 min, 70% 1 min, distilled H₂O 2 times for 1 min each).
5. Proceed to staining protocol.

ARRAY MAP



STORAGE AND STABILITY

Block	5 years
Slide	1 year

*FFPE sample stability is not universally established and is biomarker and assay dependent. More frequent re-testing may be needed for certain labile biomolecules.



Acepix Biosciences, Inc.



32990 Alvarado-Niles Road, Suite 910 Union City · CA · 94587 · USA
Tel: 1 (888) 866-0580 · Web: www.acepixonbio.com