

Human FFPE Cancer Tissue Microarray

GENERAL INFORMATION

Product Name:	Human FFPE Cancer Tissue Micro Array
Reference Number:	7300-9010 Block 7300-9020 Slide (4µm)
Date of Manufacturing:	See Product Label
Intended Use:	For IHC Control

DESCRIPTION

Tissue of Origin:	Human
Tissues Info:	See Array Map
Fixation Condition:	10% neutral buffered formalin (NBF) for 24 hours at 24-27°C

STORAGE AND STABILITY

Storage Condition	Stability*
Ambient temperature	6 months
2-8 °C with desiccation (Recommended)	2 years
-20 °C to -10 °C	5 years

*Guideline for general applications, such as immunohistochemistry (IHC) or DNA in situ hybridization (ISH). Certain biomolecules may be less stable during storage.

SAFETY AND PRECAUTIONS

This product does not contain hazardous material. Wear appropriate personal protective equipment (PPE) when handling reagents and biological specimens.

RECOMMENDED PROCEDURES

Staining using FFPE slides:


1. Bake slides at 60°C for 30-60 min.
2. Deparaffinize two times in Xylene or Xylene substitute for 5 min each time.
3. Rinse two times with 100% ethanol for 1 min each time.
4. Air dry slides or rehydrate in ethanol series (95% 2 min, 70%, 2 min, 50% 2 min, 1XPBS 2 min).
5. Proceed to staining protocol.



Acepix Biosciences, Inc.

32990 Alvarado-Niles Road Ste 910
Union City · CA · 94587 · USA
Tel: 1 (888) 866-0580
Web: www.acepibio.com

TISSUE ARRAY MAP



	1	2	3	4	5	6	7	8
A	Hs Breast Cancer	Hs Melanoma Cancer	Hs Lung Cancer	Hs Colon Cancer	Hs Normal Tonsil	Hs Liver Cancer	Hs Prostate Cancer	Hs Brain Cancer
B			9			10		
			Hs Stomach Cancer			Hs Lymphoma Cancer		